



Title:

Certificate of Analysis (CoA)

Bacterial Endotoxins Test (ABI- TEST 0012)

Date: 4/12/2026
Date Tested: 4/10/2026
Customer: The Good Gut Co
Testing material: GHK-Cu
Lot Number: unspecified
BT Sample ID: 005000039570416
Labeled Peptide
Content/Potency: 50 mg
Storage: R.T.
Visual Description: Small clear vial: purple sample, white label, silver crimp, dark blue plastic cap.
Labeled as: GHK-Cu
Manufacturer: The Good Gut Co
Testing Purpose: (ABI-TEST 0012) test quantifies bacterial endotoxin levels to confirm that the product does not elicit pyrogenic reactions, maintaining patient safety and product integrity.



Test	Method	Specification	Result
Bacterial Endotoxins Test	USP <85>	40 EU/vial reconstituted	PASS

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

Bacterial Endotoxins Test GHK-Cu Lot unspecified

Sample ID	Description	Volume (mL)	Endotoxin Limit (EU/vial)	Lysate Sensitivity (λ , EU/mL)	Result
005000039570416	GHK-Cu	3	40 EU/vial	0.25	30.2 EU/vial

The sample passed a valid gel clot endotoxin test based on the type, mode of administration and dosage.



Andrea Castro, AS
Scientist-II
BTLabs



Verna Zheng, AS
Scientist-II
BTLabs

5730 Corporate Way | Suite 220 | West Palm Beach, FL 33407
Phone: (561) 625-0133

E-mail: info@btlabtesting.com | Website: <https://btlabtesting.com>