



Title:

## Certificate of Analysis (CoA)

### Bacterial Endotoxins Test (ABI- TEST 0012)

**Date:** 4/12/2026  
**Date Tested:** 4/10/2026  
**Customer:** The Good Gut Co  
**Testing material:** KPV  
**Lot Number:** unspecified  
**BT Sample ID:** 005000039570422



**Labeled Peptide**  
**Content/Potency:** 10 mg  
**Storage:** R.T.  
**Visual Description:** Small clear vial: white sample, white label, silver crimp, green plastic cap.  
**Labeled as:** KPV  
**Manufacturer:** The Good Gut Co  
**Testing Purpose:** (ABI-TEST 0012) test quantifies bacterial endotoxin levels to confirm that the product does not elicit pyrogenic reactions, maintaining patient safety and product integrity.

Test	Method	Specification	Result
<b>Bacterial Endotoxins Test</b>	USP <85>	40 EU/vial reconstituted	PASS

The results of the CoA relate only to the item(s) tested and applied to the sample as received.

### Bacterial Endotoxins Test KPV Lot unspecified

Sample ID	Description	Volume (mL)	Endotoxin Limit (EU/vial)	Lysate Sensitivity ( $\lambda$ , EU/mL)	Result
005000039570422	KPV	3	40 EU/vial	0.25	13.3 EU/vial

The sample passed a valid gel clot endotoxin test based on the type, mode of administration and dosage.



Andrea Castro, AS  
 Scientist-II  
 BTLabs



Verna Zheng, AS  
 Scientist-II  
 BTLabs

5730 Corporate Way | Suite 220 | West Palm Beach, FL 33407  
 Phone: (561) 625-0133

E-mail: [info@btlabtesting.com](mailto:info@btlabtesting.com) | Website: <https://btlabtesting.com>